



Test Report: GSM60B5-P1J

60W AC-DC Single Output Medical Adaptor

■ DESIGN VERIFY TEST

Output Function Test

Input Function Test

Protection Function Test

Control Function Test

Component Stress Test

■ SAFETY & E.M.C. TEST

Safety Test

E.M.C. Test

■ RELIABILITY TEST

ENVIRONMENT TEST

DESIGN VERIFY TEST

OUTPUT FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	RIPPLE & NOISE	V1 : 80 mVp-p (Max)	I/P : 230VAC O/P : FULL LOAD Ta : 25°C	V1 : 41.0 mVp-p (Max)	P
2	OUTPUT VOLTAGE TOLERANCE	V1 : -5 %~ +5 % (Max)	I/P : 80 VAC / 264 VAC O/P : FULL/ MIN LOAD Ta : 25°C	V1 : -2.00 %~ 2.24 %	P
3	LINE REGULATION	V1 : -1 %~ +1 % (Max)	I/P : 100 VAC ~ 264 VAC O/P : FULL LOAD Ta : 25°C	V1 : 0 %~ 0.12 %	P
4	LOAD REGULATION	V1 : -5 %~ +5 % (Max)	I/P : 230 VAC O/P : FULL ~MIN LOAD Ta : 25°C	V1 : -2.00 %~ 2.24 %	P
5	SET UP TIME	230VAC : 1000 ms (Max) 115VAC : 1500 ms(Max)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	230VAC/ 422 ms 115VAC/ 678 ms	P
6	RISE TIME	230VAC : 30 ms (Max) 115VAC : 30 ms (Max)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	230VAC/ 5.2 ms 115VAC/ 11.9 ms	P
7	HOLD UP TIME	230VAC : 50 ms (TYP) 115VAC : 15 ms (TYP)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	230VAC/ 144 ms 115VAC/ 25.6 ms	P
8	OVER/UNDERSHOOT TEST	< ±5%	I/P : 230 VAC O/P : FULL LOAD Ta : 25°C	TEST : 0 %	P
9	DYNAMIC LOAD	V1 : 1000 mVp-p	I/P : 230 VAC (1).O/P : FULL /Min LOAD 90%DUTY/ 1KHZ (2).O/P : FULL /Min LOAD 90%DUTY/ 3KHZ (3).O/P : FULL /Min LOAD 90%DUTY/ 5KHZ (4).O/P : FULL /Min LOAD 50%DUTY/ 120HZ Ta : 25°C	(1) 680 mVp-p (2) 652 mVp-p (3) 652 mVp-p (4) 674 mVp-p	P

INPUT FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	INPUT VOLTAGE RANGE	80VAC~264 VAC	I/P : TESTING O/P : FULL LOAD Ta : 25°C I/P : LOW-LINE-3V= 77 V HIGH-LINE+15%=300 V O/P : FULL/MIN LOAD ON : 30 Sec . OFF : 30 Sec 10MIN (AC POWER ON/OFF NO DAMAGE)	71.2 V~264V TEST : OK	P
2	INPUT FREQUENCY RANGE	47HZ ~63 HZ NO DAMAGE OSC	I/P : 80 VAC ~ 264 VAC O/P : FULL~MIN LOAD Ta : 25°C	TEST : OK	P
3	EFFICIENCY	81.5% (TYP)	I/P : 230 VAC O/P : FULL LOAD Ta : 25°C	85.87 %	P
4	INPUT CURRENT	230V/ 1 A (TYP) 115V/ 1.4 A (TYP)	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	I = 0.29 A/ 230 VAC I = 0.52 A/ 115 VAC	P
5	INRUSH CURRENT	230V/ 60 A (TYP) 115V/ 30 A (TYP) COLD START	I/P : 230 VAC I/P : 115 VAC O/P : FULL LOAD Ta : 25°C	I = 41.523 A/ 230 VAC I = 20.875 A/ 230 VAC	P
6	LEAKAGE CURRENT	50 uA / 264VAC	I/P : 264 VAC O/P : Min LOAD Ta : 25°C	FOR PATIENT 34.4 μA	P
7	NO LOAD CONSUMPTION PS-ON SHORT	< 0.1 W	I/P : 240VAC O/P : NO LOAD Ta : 25°C	< 0.0499 W	P

PROTECTION FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	OVER LOAD PROTECTION	105 % ~160 %	I/P : 230 VAC I/P : 115 VAC O/P : TESTING Ta : 25°C	128.8 %/ 230 VAC 125.9 %/ 115 VAC Hiccup Mode	P
2	OVER VOLTAGE PROTECTION	CH1 : 5.25 V ~6.75 V	I/P : 230 VAC I/P : 115 VAC O/P : MIN LOAD Ta : 25°C	6.12 V/ 230 VAC 6.15 V/ 115 VAC Shut down Re- power ON	P

3	OVER TEMPERATURE PROTECTION	SPEC : RTH2>70℃ NO DAMAGE	I/P : 230 VAC O/P : FULL LOAD	O.T.P. Active Shut down Re-power ON	P
4	SHORT PROTECTION	SHORT EVERY OUTPUT 1 HOUR NO DAMAGE	I/P : 264 VAC O/P : FULL LOAD Ta : 25℃	NO DAMAGE Hiccup Mode	P

CONTROL FUNCTION TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	ERP STEP2 COMPLIANT	LEVEL $V \geq 81.7\%$	I/P: 230 VAC/115VAC O/P:100/75/50/25/0% LOAD Ta:25℃	230V 82.863% 115V 82.727%	P

COMPONENT STRESS TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	Power Transistor (D to S) or (C to E) Peak Voltage	Q1 Rated : 700 V 10 A	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on (2) Output Short (3)Full load continue Ta : 25℃	(1) 568 V (2) 344 V (3) 564 V	P
2	Diode Peak Voltage	D100 Rated : 45 V 40 A	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on (2)Output Short (3)Full load continue Ta : 25℃	(1) 30.4 V (2) 24.7 V (3) 30.6 V	P
3	Input Capacitor Voltage	C 5 Rated : 120u /400V/105℃	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on /Off (2) Min load Turn on /Off (3)Full Load /Min load Change Ta : 25℃	(1) 344 V (2) 346 V (3) 348 V	P
4	Control IC Voltage Test	U 1 Rated : 28 V	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on /Off (2) Min load Turn on /Off (3)Full Load /Min load Change Ta : 25℃	(1) 21.1 V (2) 16.8 V (3) 20.1 V	P
5	CLAMP DIODE	D 1 Rated : 800 V 2 A	I/P : High-Line +3V = 267 V O/P : (1)Full Load Turn on (2) Output Short (3)Full load continue Ta : 25℃	(3) 480 V (4) 448 V (3) 474 V	P

■ SAFETY & E.M.C. TEST

SAFETY TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	WITHSTAND VOLTAGE	I/P-O/P : 4 KVAC/min	I/P-O/P : 4.2KVAC/min Ta : 25°C	I/P-O/P : 1.706 mA NO DAMAGE	P
2	ISOLATION RESISTANCE	I/P-O/P : 500VDC>100MΩ	I/P-O/P : 500 VDC Ta : 25°C/70%RH	I/P-O/P : 9999 MΩ NO DAMAGE	P

E.M.C TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT
1	HARMONIC	EN61000-3-2 CLASS A	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	PASS	P
2	CONDUCTION	EN55011 CLASS B	I/P : 230 VAC (50HZ) O/P : FULL/50% LOAD Ta : 25°C	PASS Test by certified Lab	P
3	RADIATION	EN55011 CLASS B	I/P : 230 VAC (50HZ) O/P : FULL LOAD Ta : 25°C	PASS Test by certified Lab	P
4	E.S.D	EN61000-4-2 AIR : 15KV / Contact : 8KV	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	CRITERIA A	P
5	E.F.T	EN61000-4-4 INPUT: 2KV	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	CRITERIA A	P
6	SURGE	IEC61000-4-5 L-N : 1KV	I/P : 230 VAC/50HZ O/P : FULL LOAD Ta : 25°C	CRITERIA A	P
7	Test by certified Lab & Test Report Prepare				

RELIABILITY TEST

ENVIRONMENT TEST

NO	TEST ITEM	SPECIFICATION	TEST CONDITION	RESULT	VERDICT																																																																
1	TEMPERATURE RISE TEST	MODEL : GSM60B12-P1J 1. ROOM AMBIENT BURN-IN : 1 HRS I/P : 230VAC O/P : FULL LOAD Ta=22℃ 2. HIGH AMBIENT BURN-IN : 2HRS I/P : 230VAC O/P : FULL LOAD Ta=42℃ <table><tr><th>NO</th><th>Position</th><th>ROOM AMBIENT Ta=22.0℃</th><th>HIGH AMBIENT Ta=42.0℃</th></tr><tr><td>1</td><td>LF1</td><td>52.9℃</td><td>68.7℃</td></tr><tr><td>2</td><td>LF2</td><td>61.8℃</td><td>76.8℃</td></tr><tr><td>3</td><td>BD1</td><td>64.6℃</td><td>79.4℃</td></tr><tr><td>4</td><td>C5</td><td>60.3℃</td><td>75.5℃</td></tr><tr><td>5</td><td>D1</td><td>72.1℃</td><td>87.0℃</td></tr><tr><td>6</td><td>D40</td><td>65.6℃</td><td>80.8℃</td></tr><tr><td>7</td><td>RTH2</td><td>65.7℃</td><td>81.0℃</td></tr><tr><td>8</td><td>T1</td><td>69.1℃</td><td>84.1℃</td></tr><tr><td>9</td><td>C105</td><td>65.9℃</td><td>80.9℃</td></tr><tr><td>10</td><td>D100</td><td>80.7℃</td><td>95.4℃</td></tr><tr><td>11</td><td>C40</td><td>65.3℃</td><td>80.6℃</td></tr><tr><td>12</td><td>Q1</td><td>66.9℃</td><td>82.0℃</td></tr><tr><td>13</td><td>D42</td><td>53.7℃</td><td>70.0℃</td></tr><tr><td>14</td><td>U1</td><td>56.3℃</td><td>72.3℃</td></tr><tr><td>15</td><td>L1</td><td>45.1℃</td><td>61.1℃</td></tr></table>			NO	Position	ROOM AMBIENT Ta=22.0℃	HIGH AMBIENT Ta=42.0℃	1	LF1	52.9℃	68.7℃	2	LF2	61.8℃	76.8℃	3	BD1	64.6℃	79.4℃	4	C5	60.3℃	75.5℃	5	D1	72.1℃	87.0℃	6	D40	65.6℃	80.8℃	7	RTH2	65.7℃	81.0℃	8	T1	69.1℃	84.1℃	9	C105	65.9℃	80.9℃	10	D100	80.7℃	95.4℃	11	C40	65.3℃	80.6℃	12	Q1	66.9℃	82.0℃	13	D42	53.7℃	70.0℃	14	U1	56.3℃	72.3℃	15	L1	45.1℃	61.1℃	P
NO	Position	ROOM AMBIENT Ta=22.0℃	HIGH AMBIENT Ta=42.0℃																																																																		
1	LF1	52.9℃	68.7℃																																																																		
2	LF2	61.8℃	76.8℃																																																																		
3	BD1	64.6℃	79.4℃																																																																		
4	C5	60.3℃	75.5℃																																																																		
5	D1	72.1℃	87.0℃																																																																		
6	D40	65.6℃	80.8℃																																																																		
7	RTH2	65.7℃	81.0℃																																																																		
8	T1	69.1℃	84.1℃																																																																		
9	C105	65.9℃	80.9℃																																																																		
10	D100	80.7℃	95.4℃																																																																		
11	C40	65.3℃	80.6℃																																																																		
12	Q1	66.9℃	82.0℃																																																																		
13	D42	53.7℃	70.0℃																																																																		
14	U1	56.3℃	72.3℃																																																																		
15	L1	45.1℃	61.1℃																																																																		
2	OVER LOAD BURN-IN TEST	NO DAMAGE 1 HOUR (MIN)	I/P : 230 VAC O/P : 126% LOAD Ta : 25℃	TEST : OK	P																																																																
3	LOW TEMPERATURE TURN ON TEST	TURN ON AFTER 2 HOUR	I/P : 264VAC/100VAC O/P : 100 % LOAD Ta= -30℃	TEST : OK	P																																																																
4	HIGH HUMIDITY HIGH TEMPERATURE HIGH VOLTAGE TURN ON TEST	AFTER 12 HOURS IN CHAMBER ON CONTROL 40℃ NO DAMAGE	I/P : 272 VAC O/P : FULL LOAD Ta= 40℃ HUMIDITY= 95 %R.H	TEST : OK	P																																																																
5	TEMPERATURE COEFFICIENT	±0.03%/℃ (0~50℃)	I/P : 230 VAC O/P : FULL LOAD	± 0.006%/℃ (0~50℃)	P																																																																
6	STORAGE TEMPERATURE TEST	1. Thermal shock Temperature : -45℃~ +90℃ 2. Temperature change rate : 25℃ / MIN 3. Dwell time low and high temperature : 30 MIN/EACH 4. Total test cycle : 5 CYCLE 5. Input/Output condition : STATIC		OK	P																																																																

7	THERMAL SHOCK TEST	1. Thermal shock Temperature : $-30^{\circ}\text{C} \sim +60^{\circ}\text{C}$ 2. Temperature change rate : $25^{\circ}\text{C} / \text{MIN}$ 3. Dwell time low and high temperature : 30 MIN/EACH 4. Total test cycle : 10 CYCLE 5. Input/Output condition : 230VAC/Full Load AC ON/OFF TEST turn on 58sec ; turn off 2sec	OK	P
8	VIBRATION TEST	1 Carton & 1 Set (1) Waveform : Sine Wave (2) Frequency : 10~500Hz (3) Sweep Time : 12min/sweep cycle (4) Acceleration : 2G (5) Test Time : 60min in each axis (X.Y.Z) (6) Ta : 25°C	TEST : OK	P
9	CAPACITOR LIFE CYCLE	SUPPOSE C105 IS THE MOST CRITICAL COMPONENT (1) I/P : 230VAC O/P : FULL LOAD Ta= 25°C LIFE TIME (2) I/P : 230VAC O/P : FULL LOAD Ta= 40°C LIFE TIME (3) I/P : 230VAC O/P : 75% LOAD Ta= 40°C LIFE TIME (4) I/P : 230VAC O/P : 50% LOAD Ta= 40°C LIFE TIME	(1) 139682HRS (2) 60746HRS (3) 96096HRS (4) 167252HRS	P
10	MTBF	MIL-HDBK-217F NOTICE S2 PARTS COUNT TOTAL FAILURE RATE : 720KHRS		P
11	DMTBF/Accelerated Life Test	Demonstration Mean Time Between Failure (Expected Life): Above 50,000 hours @ TA 50°C		P

SAMPLE	TEST RESULT	TESTER	APPROVAL
PRODUCT SAMPLE	PASS	XUJ	WANGDZ